


HI 736

Phosphorus Ultra Low Range



HANNA
instruments
www.hannainst.com

Dear Customer,

Thank you for choosing a Hanna Instruments Product. Please read this instruction manual carefully before using the instrument. If you need additional technical information, do not hesitate to e-mail us at tech@hannainst.com.

Preliminary examination:

Please examine this product carefully. Make sure that the instrument is not damaged. If any damage occurred during shipment, please notify your Dealer.

Each HI 736 meter is supplied complete with:

- Two Sample Cuvettes and Caps
- Six powder reagents for Phosphorus
- 1 x 1.5V AAA Battery
- Instruction Manual




For more details about spare parts and accessories see "Accessories".

Technical specifications:

Range	0 to 200 ppb
Resolution	1 ppb
Accuracy	±5 ppb ±5% of reading @ 25°C
Typical EMC Dev.	±5 ppb
Light Source	Light Emitting Diode @ 525 nm
Light Detector	Silicon Photocell
Method	Adaptation of the <i>Standard Methods for the Examination of Water and Wastewater, 20th edition</i> , Ascorbic Acid method. The reaction between phosphorus and the reagent causes a blue tint in the sample.
Environment	0 to 50°C (32 to 122°F); max 95% RH non-condensing
Battery Type	1 x 1.5V AAA
Auto-Shut off	After 3 minutes of non-use and 2 minutes after reading.
Dimensions	81.5 x 61 x 37.5 mm (3.2 x 2.4 x 1.5")
Weight	64 g (2.25 oz.).

Functional description:



1. Cuvette cap.
2. Cuvette with cap.
3. Cuvette holder.
4. Liquid Crystal Display.
5. Button

Errors and warnings:

L.H.

Light High: There is too much light to perform a measurement. Please check the preparation of the zero cuvette.

L.Lo

Light Low: There is not enough light to perform a measurement. Please check the preparation of the zero cuvette.

Inu

Inverted cuvettes: The sample and the zero cuvette are inverted.

0

Under range: A blinking "0" indicates that the sample absorbs less light than the zero reference. Check the procedure and make sure you use the same cuvette for reference (zero) and measurement.

200

Over Range: A flashing value of the maximum concentration indicates an over range condition. The concentration of the sample is beyond the programmed range: dilute the sample and re-run the test.

bAt

Battery low: The battery must be replaced soon.


bAd

Dead battery: This indicates that the battery is dead and must be replaced. Once this indication is displayed, normal operation of the instrument will be interrupted. Change the battery and restart the meter.


bAt

Measurement procedure:

- Turn the meter on by pressing the button. After all the segments are displayed, "C.1", "Add" appears with "Press" blinking, the meter is ready.




- Fill the cuvette with 10 mL of unreacted sample and replace the cap. Place the cuvette into the meter and close the meter's cap.



- Press the button. When the display shows "Add", "C.2" with "Press" blinking the meter is zeroed.



- Remove the cuvette, open it and add the content of one packet of HI 736-25 reagent. Replace the cap and shake gently for 2 minutes until the powder is completely dissolved. Replace the cuvette into the meter.



- Press and hold the button until the timer is displayed on the LCD.



- The instrument directly displays the concentration of phosphorus in ppb. The meter automatically turns off after 2 minutes.



Tips for an accurate measurement

- It is important that the sample does not contain any debris.
- Whenever the cuvette is placed into the measurement cell, it must be dry outside, and completely free of fingerprints, oil or dirt. Wipe it thoroughly with HI 731318 or a lint-free cloth prior to insertion.
- Shaking the cuvette can generate bubbles, causing higher readings. To obtain accurate measurements, remove such bubbles by swirling or by gently tapping the cuvette.
- Do not let the reacted sample stand for too long after reagent is added, or accuracy will be lost.
- After the reading it is important to immediately discard the sample, otherwise the glass might become permanently stained.

Battery management

To save the battery, the instrument shuts down after 3 minutes of non-use and 2 minutes after read.


One fresh battery lasts for a minimum of 5000 measurements, depending on the light level. When the battery capacity is under 10 % "bAt" appears on the LCD at start-up.

If the battery is empty and accurate measurements can't be taken any more, the instrument shows "bAd" then "bAt" each for 1 second and turns off.

To restart the instrument, the battery must be replaced with a fresh one.

To replace the instrument's battery, follow the steps:

- Turn the instrument off by holding the button until the meter shuts off.
- Turn the instrument upside down and remove the battery cover with a screwdriver.



- Remove the battery from its location and replace it with a fresh one.
- Insert the battery cover and replace the screw with a screwdriver.

Accessories:

REAGENT SETS

HI 736-25 Reagents for 25 Phosphorus Ultra Low Range tests

OTHER ACCESSORIES

HI 740028 1.5V AAA batteries (4 pcs)
HI 731318 Tissue for wiping cuvettes (4 pcs)
HI 731321 Glass cuvettes (4 pcs)
HI 731353 Caps for cuvettes (4 pcs)
HI 93703-50 Cuvettes cleaning solution (230 mL).

Recommendations for Users

Before using these products, make sure that they are entirely suitable for your specific application and for the environment in which they are used.

Operation of these instruments may cause unacceptable interferences to other electronic equipments, thus requiring the operator to take all necessary steps to correct interferences.

Any variation introduced by the user to the supplied equipment may degrade the instrument's EMC performance.

To avoid damages or burns, do not put the instrument in microwave oven. For the safety of you and the instrument do not use or store the instrument in hazardous environments.

Hanna Instruments reserves the right to modify the design, construction and appearance of its products without advance notice.

For additional information, contact your dealer or the nearest Hanna Customer Service Center. To find the Hanna Office in your area, visit our web site

www.hannainst.com

HANNA[®]
instruments